















Features

- Universal AC input / Full range
- 2 pole AC inlet IEC320-C8
- Medical safety approved (2 x MOPP between primary to secondary)
- Suitable for BF application with appropriate system consideration
- Low leakage current <50μA
- No load power consumption<0.1W
- Energy efficiency Level VI (except 5~9V for Level V)
- Comply with EISA 2007/DoE,NRCan,AU/NZ MEPS,EU ErP and meet CoC Version 5
- High efficiency up to 88%
- High operating temperature up to +60°C
- Class II power (without earth pin)
- · Protections: Short circuit / Overload / Over voltage
- · Fully enclosed plastic case
- · LED indicator for power on
- · Optional lock type DC plug
- · 100% full load burn-in test
- · 3 years warranty

Applications


- · Blood glucose meter
- Blood pressure meter
- Nebulizer
- Inhaler
- · Portable medical device

Description

GSM36B is a highly reliable, 36W desktop style single-output green medical adaptor series. This product is equipped with a 2-pin (no FG) standard IEC320-C8 power plug, adopting the input range from 80VAC to 264VAC. The entire series supplies different output voltages between 5VDC and 48VDC that can satisfy the demands for various kinds of miniature medical devices. The circuitry design meets the international medical standards ($2 \times MOPP$), having an ultra low leakage current ($<50\mu A$), fitting the medical devices in direct electrical contact with the patients.

With the efficiency up to 88% and the extreme low no-load power consumption below 0.1W, GSM36B is compliant with USA EISA 2007/DoE, Canada NRCan, Australia and New Zealand MEPS, EU Erp and meet Code of Conduct(CoC) Version 5; the supreme feature allows the adaptor to save the energy when it is either under the operating mode or the standby mode. The entire series utilizes the 94V-0 flame retardant plastic case, providing the double insulation that effectively prevents electrical shock. GSM36B is approved with the international medical safety certificates.

Model Encoding



36W AC-DC High Reliability Medical Adaptor

SPECIFICATION

| ORDER NO. | | GSM36B05-P1J | GSM36B07-P1J | GSM36B09-P1J | GSM36B12-P1J | GSM36B15-P1J | GSM36B18-P1J | GSM36B24-P1J | GSM36B48-P1J |
|-------------|--|--|-------------------|------------------|--------------------|---------------------|--------------|--------------|--------------|
| ОИТРИТ | SAFETY MODEL NO. | GSM36B05 | GSM36B07 | GSM36B09 | GSM36B12 | GSM36B15 | GSM36B18 | GSM36B24 | GSM36B48 |
| | DC VOLTAGE Note.2 | 5V | 7.5V | 9V | 12V | 15V | 18V | 24V | 48V |
| | RATED CURRENT | 4.5A | 4.32A | 4A | 3A | 2.4A | 2A | 1.5A | 0.75A |
| | CURRENT RANGE | 0 ~ 4.5A | 0 ~ 4.32A | 0 ~ 4A | 0 ~ 3A | 0 ~ 2.4A | 0 ~ 2A | 0 ~ 1.5A | 0 ~ 0.75A |
| | RATED POWER (max.) | 22.5W | 32.4W | 36W | 36W | 36W | 36W | 36W | 36W |
| | RIPPLE & NOISE (max.) Note.3 | 80mVp-p | 80mVp-p | 80mVp-p | 120mVp-p | 120mVp-p | 150mVp-p | 180mVp-p | 240mVp-p |
| | VOLTAGE TOLERANCE Note.4 | ±6.0% | ±5.0% | ±5.0% | ±3.0% | ±3.0% | ±3.0% | ±2.0% | ±2.0% |
| | LINE REGULATION Note.5 | ±1.0% | ±1.0% | ±1.0% | ±1.0% | ±1.0% | ±1.0% | ±1.0% | ±1.0% |
| | LOAD REGULATION | ±6.0% | ±5.0% | ±5.0% | ±3.0% | ±3.0% | ±3.0% | ±2.0% | ±2.0% |
| | SETUP, RISE TIME Note.6 | 500ms, 30ms / 2 | 230VAC 10 | 00ms, 30ms / 11 | VAC at full load | | | | |
| | HOLD UP TIME (Typ.) | 16ms / 230VAC 16ms / 115VAC at full load | | | | | | | |
| | VOLTAGE RANGE Note.7 | 80 ~ 264VAC 113 ~ 370VDC | | | | | | | |
| INPUT | FREQUENCY RANGE | 47 ~ 63Hz | | | | | | | |
| | EFFICIENCY (Typ.) | 80% | 83% | 84% | 86% | 87% | 87% | 87% | 88% |
| | AC CURRENT (Typ.) | 0.9A / 115VAC | | | | | | | |
| | INRUSH CURRENT (Typ.) | 55A / 230VAC 30A / 115VAC | | | | | | | |
| | LEAKAGE CURRENT(max.) | Touch current < 50µA/264VAC | | | | | | | |
| PROTECTION | OVERLOAD | 105 ~ 170% rated output power | | | | | | | |
| | | Protection type | : Hiccup mode, | recovers autom | atically after fau | ılt condition is re | moved | | |
| | OVER VOLTAGE | 5.25 ~ 7.5V | 7.88 ~ 10.5V | 9.45 ~ 13V | 12.6 ~ 17.2V | 15.75 ~ 20.25V | 18.9 ~ 25.2V | 25.2 ~ 32.4V | 50.4 ~ 64.8V |
| | | Protection type | : Shut down o/p | voltage, re-pov | er on to recover | r | | | |
| ENVIRONMENT | WORKING TEMP. | -25 ~ +60°C (Refer to "Derating Curve") | | | | | | | |
| | WORKING HUMIDITY | 20% ~ 90% RH non-condensing | | | | | | | |
| | STORAGE TEMP., HUMIDITY | -40 ~ +85°C, 10 ~ 95% RH | | | | | | | |
| | TEMP. COEFFICIENT | ±0.03% / °C (0~40°C) | | | | | | | |
| | VIBRATION | 10 ~ 500Hz, 2G 10min./1cycle, period for 60min. each along X, Y, Z axes | | | | | | | |
| | SAFETY STANDARDS | ANSI/AAMI ES60601-1 / 60601-1-11, TUV EN60601-1 / 60601-1-11, EAC TP TC 004 approved | | | | | | | |
| SAFETY & | WITHSTAND VOLTAGE | I/P-O/P:4KVAC | | | | | | | |
| EMC | ISOLATION RESISTANCE | I/P-O/P:100M Ohms / 500VDC / 25°C / 70% RH | | | | | | | |
| (Note. 8) | EMC EMISSION | Compliance to EN55011(CISPR11) class B, EN61000-3-2,3, FCC PART 15 class B, EAC TP TC 020 | | | | | | | |
| | EMC IMMUNITY | Compliance to EN61000-4-2,3,4,5,6,8,11, EN55024, EN60601-1-2, EN61204-3 medical level, criteria A, EAC TP TC 020 | | | | | | | |
| OTHERS | MTBF | 657.4K hrs min. MIL-HDBK-217F(25 $^{\circ}$ C) | | | | | | | |
| | DIMENSION | 79*54*33mm (L*W*H) | | | | | | | |
| | PACKING | 235g; 60pcs / 1 | 5.1Kg / CARTON | ٧ | | | | | |
| CONNECTOR | PLUG | See page 3; Ot | her type availabl | e by customer re | quested | | | | |
| | CABLE | See page 3; Ot | her type availabl | e by customer re | quested | | | | |
| NOTE | All parameters are specified at 230VAC input, rated load, 25°C 70% RH ambient. DC voltage: The output voltage set at point measure by plug terminal & 50% load. Ripple & noise are measured at 20MHz by using a 12" twisted pair terminated with a 0.1uf & 47uf capacitor. Tolerance: includes set up tolerance, line regulation, load regulation. Line regulation is measured from low line to high line at rated load. Length of set up time is measured at first cold start. Turning ON/OFF the power supply may lead to increase of the set up time. Derating may be needed under low input voltage. Please check the derating curve for more details. The power supply is considered as an independent unit, but the final equipment still need to re-confirm that the whole system complies with the EMC directives. For guidance on how to perform these EMC tests, please refer to "EMI testing of component power supplies." (as available on http://www.meanwell.com) | | | | | | | | |



